

Mixed Analyte Performance Evaluation Program

Statistical Summary

Sample ID: MAPEP-06-GrF15

Gross alpha / beta Radiological Air Filters

Analyte	T(1)	A(2)	Grand Mean	Std. Dev.	Reference Value	Analyte Text	Acceptance Limits	Units
Gross alpha	57	56	0.334	0.060	0.361		>0.0 - 0.722	(Bq/sample)
Gross beta	58	52	0.425	0.052	0.481		0.241 - 0.722	(Bq/sample)

Gross Alpha Flags:

A = Result acceptable, Bias \leq +/- 100% with a statistically positive result at two standard deviations (Result/Uncertainty > 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

N = Result not acceptable, Bias $>$ +/- 100% or the reported result is not statistically positive at two standard deviations (Result/Uncertainty ≤ 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, includes zero).

Gross Beta Flags:

A = Result acceptable, Bias \leq +/- 50% with a statistically positive result at two standard deviations (Result/Uncertainty > 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

N = Result not acceptable, Bias $>$ +/- 50% or the reported result is not statistically positive at two standard deviations (Result/Uncertainty ≤ 2 , i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, includes zero).

Note:

Outliers are excluded from the statistical summary.

Outliers are defined as laboratory data with a bias greater than 60 percent for gamma spectrometry analyses and 30 percent for all others.

(1) T = Total Number of Laboratories Reporting Analyte.

(2) A = Number of Laboratories with 'Acceptable' Performance.